



DRAFT - URGENT VOLUNTARY MEDICAL DEVICE RECALL

Boston Conditioning Solution Original Formula

Date

Dear Bausch & Lomb Customer:

Bausch & Lomb is conducting a voluntary recall of one lot of Boston Conditioning Solution Original Formula because it failed to meet one of the shelf life specifications at the end of the expiry period. This action affects only bottle lot number GA7060. No other lots of Boston Conditioning Solution Original Formula are affected.

This one lot of Boston Conditioning Solution Original Formula was packaged into three different care kit carton lots. This action only affects those starter kit carton lots containing Boston Conditioning Solution Original Formula (1 OZ) lot number GA7060. See the below table for details.

Bottle Lot #	Carton Lot #	Size	Item #	Expiration Date
GA7060	AD7043; AE7004; AE7083	1 OZ	7474G	1/2009

Our records show that we shipped a quantity of the above product lot to you between May and July, 2007. Based on normal usage, we expect that you do not have any of this product remaining in your inventory. However, as a precaution we ask that you please take the following steps:

1. Check your inventory for the above Boston Conditioning Solution Original Formula and segregate them for return shipment. The affected product can be most easily identified by checking the lot number on the outside carton (on the right side of the back panel) to see if it matches the lot numbers provided above.
2. Complete the enclosed Business Reply Card and return it via US mail within 5 days. **It is important to return the Card even if you do not have any of the Boston Conditioning Solution Original Formula identified above.**
3. Complete the enclosed Packing Slip and affix the provided pre-paid UPS shipping label to return your product to Stericycle Inc., 2670 Executive Drive, Suite A, Indianapolis, IN 46241; the firm conducting this product recall on behalf of Bausch & Lomb.

If you have questions about the recall process or need additional shipping labels, please contact Stericycle Inc. at 1.877.546.0481. If you or your patients have any questions regarding this recall, please contact Bausch & Lomb's Customer Resource Center at 1.800.828.9030. We appreciate your prompt assistance and apologize for any inconvenience this may have caused.

Sincerely,

Michael A. Santalucia